



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 21, 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL

DISTRICT: 6 ADDRESS: SR28 MP13

LOCATION: Rock Is 94 COUNTY: Douglas

MAKE: GSE TYPE: Trk Scale ACCURACY CLASS: IIII

MODEL: 651 ID. NO: 201450 CAPACITY: 200000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

Dual


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
PLATFORM 1				PLATFORM 2			
SEC 1	20,000	20,000	0	SEC 1	20,000	20,000	0
CENTER	20,000	20,000	0	CENTER	20,000	20,000	0
SEC 2	20,000	20,000	0	SEC 2	20,000	20,000	0

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-1 and 200446-0-L3184-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: June 14, 2012

Print name: David Cromer



State of Washington

County of Thurston

Signed and sworn to (or affirmed before me on JUNE 14, 2012)

Patricia A. Geppert

(Signature of notary public)

PATRICIA A. GEPPERT

Printed Name of Notary

My appointment expires SEPT. 4, 2015